

NRFit and luer lock syringe composition comparison



Brand	Туре	Name	Product Code	Total Graduated Capacity (mL)	Barrel	Plunger Rod	Gasket/ Grommet Plunger Stopper	Gasket/Grommet/ Plunger Stopper Vulcanising Agent	Lubricant	Sterilisation Method	Notes
BBraun	NRFit	Omnifix Lock NRFit	4617003	3	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
BBraun	NRFit	Omnifix Lock NRFit	4617006	5	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	NRFit	Omnifix Lock NRFit	4617014	10	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	NRFit	Omnifix Lock NRFit	4617021	20	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	NRFit	Omnifix Lock NRFit	4617508-01	50	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	Luer Lock	Omnifix Luer Lock Solo	4617029V	2	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	Luer Lock	Omnifix Luer Lock Solo	4617022V	3	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	Luer Lock	Omnifix Luer Lock Solo	4617053V	5	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	Luer Lock	Omnifix Luer Lock Solo	4617100V	10	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	Luer Lock	Omnifix Luer Lock Solo	4617207V	20	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	Luer Lock	Omnifix Luer Lock Solo	4617304F	30	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
Braun	Luer Lock	Omnifix Luer Lock Solo	4617509F	50	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil mixture	Ethylene oxide	
BD	NRFit	BD Syringe NRFit Lok	400050	3	Polypropylene	Polypropylene	Polyisoprene Rubber	No data available	Medical grade Silicone	Gamma or E-beam Irradiation	The narrower tip of BD® Syringes NRFit™ Lok and NRFit™ Slip have reduced dead space and drug waste compared to BD Luer-Lok™ and Luer Slip Tip Syringes. From a Physicochemical stability perspectiuve this is preferable. BD provided statement that BD Luer Lok syringe code 309658 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.
BD	NRFit	BD Syringe NRFit Lok	400173	5	Polypropylene	Polypropylene	Polyisoprene Rubber	No data available	Medical grade Silicone	Gamma or E-beam Irradiation	The narrower tip of BD® Syringes NRFit™ Lok and NRFit™ Slip have reduced dead space and drug waste compared to BD Luer-Lok™ and Luer Slip Tip Syringes. From a Physicochemical stability perspectitive this is preferable. BD provided statement that BD Luer Lok syringe code 309649 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.
BD	NRFit	BD Syringe NRFit Lok	400174	10	Polypropylene	Polypropylene	Polyisoprene Rubber	No data available	Medical grade Silicone	Gamma or E-beam Irradiation	The narrower tip of BD® Syringes NRFit™ Lok and NRFit™ Slip have reduced dead space and drug waste compared to BD Luer-Lok™ and Luer Slip Tip Syringes. From a Physicochemical stability perspective this is preferable. BD provided statement that BD Luer Lok syringe code 300912 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.
BD	NRFit	BD Syringe NRFit Lok	400182	20	Polypropylene	Polypropylene	Polyisoprene Rubber	No data available	Medical grade Silicone	Gamma Irradiation	The narrower tip of BD® Syringes NRFit™ Lok and NRFit™ Slip have reduced dead space and drug waste compared to BD Luer-Lok™ and Luer Slip Tip Syringes. From a Physicochemical stability perspective this is preferable. BD provided statement that BD Luer Lok syringe code 302830 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.
3D	NRFit	BD Syringe NRFit Lok	400183	50	Polypropylene	Polypropylene	Polyisoprene Rubber	No data available	Medical grade Silicone	Gamma Irradiation	The narrower tip of BD® Syringes NRFit™ Lok and NRFit™ Slip have reduced dead space and drug waste compared to BD Luer-Lok™ and Luer Slip Tip Syringes. From a Physicochemical stability perspective this is preferable. BD provided statement that BD Luer Lok syringe code 309653 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.
3D	Luer Lock	BD Hypodermic Syringe Luer-Lok	309628	1	Polypropylene with polycarbonate tip	Polypropylene	Natural Rubber/ Synthetic Rubber/ Thermoplastic Elastomers	No data available	Medical grade Silicone	Irradiation	Manufcatured in USA
BD	Luer Lock	BD Hypodermic Syringe Luer-Lok	309658	3	Polypropylene	Polypropylene	Natural Rubber/ Synthetic Rubber/ Thermoplastic Elastomers	No data available	Medical grade Silicone	Irradiation	Manufactured in USA. BD provided statement that BD NRFit syringe code 400050 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.
BD	Luer Lock	BD Hypodermic Syringe Luer-Lok	309649	5	Polypropylene	Polypropylene	Natural Rubber/ Synthetic Rubber/ Thermoplastic Elastomers	No data available	Medical grade Silicone	Irradiation	Manufactured in USA. BD provided statement that BD NRFit syringe code 400173 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.
BD	Luer Lock	BD Hypodermic Syringe Luer-Lok	300912	10	Polypropylene	Polypropylene	Natural Rubber/ Synthetic Rubber/ Thermoplastic Elastomers	No data available	Medical grade Silicone	Irradiation	Manufactured in USA. BD provided statement that BD NRFit syringe code 400174 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.
BD	Luer Lock	BD Hypodermic Syringe Luer-Lok	302830	20	Polypropylene	Polypropylene	Natural Rubber/ Synthetic Rubber/ Thermoplastic Elastomers	No data available	Medical grade Silicone	Irradiation	Manufactured in USA. BD provided statement that BD NRFit syringe code 400182 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.
BD	Luer Lock	BD Hypodermic Syringe Luer-Lok	309653	50	Polypropylene	Polypropylene	Natural Rubber/ Synthetic Rubber/ Thermoplastic Elastomers	No data available	Medical grade Silicone	Irradiation	Manufactured in USA. BD provided statement that BD NRFit syringe code 400183 has the same design, materials of construction in fluid path and manufacturing process as this syringe. The only difference is the design of the syringe tip.



NRFit and luer lock syringe composition comparison



											Service
Brand	Туре	Name	Product Code	Total Graduated Capacity (mL)	Barrel	Plunger Rod	Gasket/ Grommet Plunger Stopper	Gasket/Grommet/ Plunger Stopper Vulcanising Agent	Lubricant	Sterilisation Method	Notes
BD	Luer Lock	BD Plastipak Syringe Luer Lok	- 305959	10	Polypropylene	Polypropylene	Polyisoprene black rubber or Black TPE Copolymer	No data available	Medical grade Silicone	Gas mixture of Ethylene Oxide and CO2 (in the proportion 90:10)	Manufactured in Spain
BD	Luer Lock	BD Plastipak Syringe Luer Lok	- 300629	20	Polypropylene	Polypropylene	Polyisoprene black rubber or Black TPE Copolymer	No data available	Medical grade Silicone	Gas mixture of Ethylene Oxide and CO2 (in the proportion 90:10)	Manufactured in Spain
BD	Luer Lock	BD Plastipak Syringe Luer Lok	- 301189	20	Polypropylene	Polypropylene	Polyisoprene black rubber or Black TPE Copolymer	No data available	Medical grade Silicone	Gas mixture of Ethylene Oxide and CO2 (in the proportion 90:10)	Manufactured in Spain
BD	Luer Lock	BD Plastipak Syringe Luer Lok	- 301229	30	Polypropylene	Polypropylene	Polyisoprene black rubber or Black TPE Copolymer	No data available	Medical grade Silicone	Gas mixture of Ethylene Oxide and CO2 (in the proportion 90:10)	Manufactured in Spain
BD	Luer Lock	BD Plastipak Syringe Luer Lok	- 300865	50	Polypropylene	Polypropylene	Polyisoprene black rubber or Black TPE Copolymer	No data available	Medical grade Silicone	Gas mixture of Ethylene Oxide and CO2 (in the proportion 90:10)	Manufactured in Spain
GBUK	NRFit	Dash 6 NRFit Lock	15011010	1	Polypropylene	Polypropylene	Isoprene rubber	No data available	Silicone oil	Ethylene oxide	
GBUK	NRFit	Dash 6 NRFit Lock	15011030	3	Polypropylene	Polypropylene	Isoprene rubber	No data available	Silicone oil	Ethylene oxide	
GBUK	NRFit	Dash 6 NRFit Lock	15011050	5	Polypropylene	Polypropylene	Isoprene rubber	No data available	Silicone oil	Ethylene oxide	
GBUK	NRFit	Dash 6 NRFit Lock	15011100	10	Polypropylene	Polypropylene	Isoprene rubber	No data available	Silicone oil	Ethylene oxide	
GBUK	NRFit	Dash 6 NRFit Lock	15011200	20	Polypropylene	Polypropylene	Isoprene rubber	No data available	Silicone oil	Ethylene oxide	
GBUK	NRFit	Dash 6 NRFit Lock	15011600	60	Polypropylene	Polypropylene	Isoprene rubber	No data available	Silicone oil	Ethylene oxide	
Terumo	Luer Lock	Concentric Luer Lock	SS*02LE1	2.5	Polypropylene	Polypropylene	Thermoplastic elastomer	No data available	Silicone oil	Radiation (e-beam)	
Terumo	Luer Lock	Concentric Luer Lock	MDSS03LE	3	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil	Ethylene oxide	
Terumo	Luer Lock		MDSS05LE	5	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil	Ethylene oxide	
Terumo	Luer Lock	Concentric Luer Lock	SS*05LE1	5	Polypropylene	Polypropylene	Thermoplastic elastomer	No data available	Silicone oil	Radiation (e-beam)	
Terumo	Luer Lock	Concentric Luer Lock	MDSS10LE	10	Polypropylene	Polypropylene	Polyisoprene	No data available	Silicone oil	Ethylene oxide	
Terumo	Luer Lock		SS*10LE1	10	Polypropylene	Polypropylene	Thermoplastic elastomer	No data available	Silicone oil	Radiation (e-beam)	
Terumo	Luer Lock	Concentric Luer Lock	SS+20L1	20	Polypropylene	Polypropylene	Thermoplastic elastomer	No data available	Silicone oil	Radiation (e-beam)	
Terumo	Luer Lock	Concentric Luer Lock	SS+30L1MP	30	Polypropylene	Polypropylene	Thermoplastic elastomer	No data available	Silicone oil	Radiation (e-beam)	
Terumo	Luer Lock	Concentric Luer Lock	SS+50L1	50	Polypropylene	Polypropylene	Thermoplastic elastomer	No data available	Silicone oil	ummary	
Vygon	NRFit	Connexion NRFit Lock	V051010032A		Polypropylene	Polypropylene	Isoprene compound	S (sulfur) 1%	Polydimethylsiloxane	,	
Vygon	NRFit	Connexion NRFit Lock	V051010052A		Polypropylene	Polypropylene	Isoprene compound	S (sulfur) 1%	Polydimethylsiloxane	,	
Vygon	NRFit	Connexion NRFit Lock	V051010102A		Polypropylene	Polypropylene	Isoprene compound	S (sulfur) 1%	Polydimethylsiloxane	,	
Vygon	NRFit	Connexion NRFit Lock	V051010202A		Polypropylene	Polypropylene	Isoprene compound	S (sulfur) 1%	Polydimethylsiloxane	,	
Vygon	NRFit	Connexion NRFit Lock	V051010602A	60	Polypropylene	Polypropylene	Isoprene compound	S (sulfur) 1%	Polydimethylsiloxane	Ethylene oxide	